trophon2®
Protection by design

Cleaning & Disinfection Guide

Ensure the trophon2 device is powered off prior to cleaning or disinfection.

Cleaning: Nanosonics recommends that the surfaces of the trophon2 device be cleaned in accordance with organisational policies and following the detection of visible contamination. When cool, wipe the chamber and outside surfaces of the trophon2 device with a cloth or wipe moistened with a mild general purpose detergent, or soap and water solution, such as dish soap or general cleaners, until all surfaces are visibly clean.

Disinfecting: Nanosonics recommends that the surfaces of the trophon2 device be disinfected in accordance with organisational policies. Wipe all accessible surfaces of the trophon2 device with an Isopropanol or Quaternary Ammonium (Quat) wipe.

Note: The inside of trophon2 chamber is subjected to a disinfecting agent during each disinfection cycle and should not require additional manual disinfection with a wipe.

Cautions: Do NOT submerge the trophon2 device, or pour liquids over it.

Keep the trophon2 device level and upright at all times.

Keep the power socket completely dry.

Outside Surfaces

CASING & SCREEN	HANDLE	CABLE TRAY
	 INSIDE & OUTSIDE	 

Inside Surfaces

CHAMBER	SILICONE SEALS	IPP AND GLAND SEAL	CABLE CLAMP
 	 	 	 

Nanosonics Limited (Manufacturer) 7-11 Talavera Road, Macquarie Park NSW 2113, Australia.
T: +61 2 8063 1600 E: info@nanosonics.com.au W: www.nanosonics.com.au

© 2022 Nanosonics Limited. All rights reserved. MM01494-GLO-GU V02 1686180. November 2022

nanosonics
Infection Prevention. For Life.